
Civil Engineering Testing Equipment

www.nl-test.com



J-RING APPARATUS For Self-Compacting Concrete (SCC) Standard: EN 12350-12

Self Compacting Concrete

Used to determine the passing ability of self-sompacting concrete to flow through obstructions, tight openings and spaces between reinforcing bars without segregation or blocking. The time required (t500) for concrete to flow to a diameter of 500 mm is also measured.

Technical Specifications:

Model Number	NL 4028 X / 001

Unit Consists Of:

Model Number	Parts Description
NL 4028 X / 001 - P001	J-Ring Apparatus
NL 4013 X / 001 - P001	Slump Cone
NL 4028 X / 001 - P002	Base Plate 905 x 905 mm (S.Steel)

^{*1} Copy of Manual Instruction

Copyright 2016 NL Scientific Instruments Sdn. Bhd. All Rights Reserved. Revision 3(All information in this catalogue only for customer's reference and not for quotation. Design and dimension maybe change without prior notice.)